



شرکت پارس تسو (سهامی خاص)

نماینده انحصاری کمپانی MIC در ایران

دستگاه اکسیژن ساز

مدل CP-501i

با ظرفیت 5 لیتر

سیستم نمایشگر دیجیتال لحظه ای خلوص اکسیژن

سیستم الارم LED و صوتی

صدای کم در حد ۴۵ دسی بل

وزن کم و ابعادی کوچک برای حمل و نقل آسان

یک سال گارانتی و ۵ سال خدمات پس از فروش

دارای گواهینامه های استاندارد کیفیت از DNV نروژ و سایر شرکت های معتبر

IEC 60601-1 , IEC 60601-1-2 , ISO 13485, ISO 8359 , CE 9342

تهران، خیابان الوند، کوچه پرمک، شماره ۱۰، واحد ۵ تلفن: ۸۸۷۷۴۵۷۵ فکس: ۸۸۶۷۸۷۷۹



**DET NORSKE VERITAS**  
**EC CERTIFICATE - FUEL QUALITY ASSURANCE SYSTEM**

Certificate No. 18903-2011-03-001-NA  
This Certificate consists of 2 pages

This is a certificate for the Quality Management System of

**Foshan MIC Medical Technology Co., Ltd.**

Unit 3, 5<sup>th</sup> Floor, No.7 Building, Zone A, Huarun Industrial Park, 17<sup>th</sup> Shekai Road, Gaibang, Nanshi, Foshan, Guangdong, China

for design, production and final product inspection of

**Electro Medical Devices**

See Annex A and Annex B for the conformity assessment procedure described in Article 6.1.a and Annex E excluding section 6.1 (Article 6.1 of Council Directive 90/269/EEC on Medical Devices, as amended, and based on (2007)

Further details are given on sheet

Place and date  
Nanshi, 17 January 2012

The Certificate is valid until  
17 January 2017

For DET NORSKE VERITAS CERTIFICATION AS  
NOTIFIED BODY

*Cecilie Guderson Torg*  
Cecilie Guderson Torg  
Certificate Manager

**CE**  
Notified Body No.  
0154

Stein Helle Nysten  
Technical Director

Public Declaration: I declare under penalty of perjury that the information provided in this certificate is true and correct. I understand that any false information provided in this certificate may result in the revocation of this certificate and may be subject to legal action.

Det Norske Veritas Certification AS  
P.O. Box 122, N-2007 Kjeller, Norway  
Tel: +47 91 12 12 12  
www.dnv.com

**ISO 13485**



**DET NORSKE VERITAS**  
**MANAGEMENT SYSTEM CERTIFICATE**

Certificate No. 18903-2011-03-001-NA

This is a certificate for the Management System of

**Foshan MIC Medical Technology Co., Ltd.**

Unit 3, 5<sup>th</sup> Floor, No.7 Building, Zone A, Huarun Industrial Park, 17<sup>th</sup> Shekai Road, Gaibang, Nanshi, Foshan, Guangdong, China

has been found to conform to the standard

**ISO 13485:2003**

This Certificate is valid for the following product or service ranges:

**Design, Manufacture and Sales of Oxygen Concentrator and Nebulizer;  
Sales of related accessories and parts.**

Issue of Certificate date  
27 December 2011

The Certificate is valid until  
27 December 2016

The activities have been performed under the  
supervision of  
Liz Jordan Esq  
Lead Auditor

Place and date  
Nanshi, 27 December 2011

Det Norske Veritas  
CERTIFICATION AS  
NOTIFIED BODY

*Eugene Wenger Hestnes*  
Eugene Wenger Hestnes  
Management Representative

Each of the signatories of this Certificate is bound by the conditions of use of this Certificate Agreement and the Certificate itself.  
The conditions of use apply to the certificate and are available at [www.dnv.com](http://www.dnv.com)  
DET NORSKE VERITAS CERTIFICATION AS  
P.O. Box 122, N-2007 Kjeller, Norway  
Tel: +47 91 12 12 12  
www.dnv.com

**CE 9342**

کیفیت تولید و عملکرد این دستگاه توسط شرکتهای معتبر بین المللی کنترل کیفیت تضمین گردیده و ضمن آنکه بالاترین راندمان تولید اکسیژن با خلوص بالای 90% را در اختیار مصرف کننده قرار میدهد، سیستم نمایشگر دیجیتالی خلوص اکسیژن نیز این امکان را به مصرف کننده میدهد که از خلوص استاندارد اکسیژن مصرفی خود اطمینان حاصل نماید.

نمایشگر خلوص اکسیژن



**Real Time Oxygen Density Display**

- تا 5 لیتر در دقیقه 90% ، تا 4 لیتر در دقیقه 93% ، تا 2 لیتر در دقیقه 94% ،  
حرارت از 10 تا 35 درجه سانتی گراد، رطوبت از 15% تا 95% فشار اتمسفر 50 تا 106 Kpa

**مشخصات دستگاه اکسیژن ساز مدل CP-501i**

- 1 - اکسیژن قابل دسترسی 1 تا 5 لیتر در دقیقه
- 2 - برق ورودی 220 ولت 50 هرتز 1.8 آمپر
- 3 - درصد اکسیژن 1 تا 5 لیتر در دقیقه 93%+3%
- 4 - وزن 16.6 کیلوگرم
- 5 - سطح صدا ISO 8359 45 دسی بل
- 6 - ابعاد 400x300 x510 میلیمتر
- 7 - دوره عمر دستگاه 5 سال
- 8 - دوره گارانتی یک سال
- 9 - مقدار متوسط اکسیژن 2 تا 4 لیتر در دقیقه 94% ، تا 5 لیتر در دقیقه 93%
- 10 - محدوده کار در محیط

- برای اطلاعات بیشتر در مورد این محصول و سایر محصولات شرکت پارس تسو می توانید به سایت ما مراجعه نمایید

Website : [www.parstaso.ir](http://www.parstaso.ir)

E-mail : [info@parstaso.ir](mailto:info@parstaso.ir)